

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 00555 Calib. Date: 03/07/2012 Calib. No.: 00015  
Certification File No.: 00556 Cert. Date: 03/07/2012 Cert. No.: 00009  
Linearity File No.: 00557 Lin. Date: 03/07/2012 Lin. No.: 00009  
Solution File No.: 00638 Soln. Date: 07/17/2012 Soln. No.: 00118  
Sequential File No.: 00638 File Date: 07/17/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113  
Control Solution %: 0.100% Expires: 08/17/2013  
Solution Control Lot: 11H094 Bottle No.: 0260

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:30D		
Control 1 EC	0.100%	11:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:31D		
Control 2 EC	0.099%	11:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:32D		
Control 3 EC	0.099%	11:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PTL K #129*

Badge No.: 129

Date: 07/17/2012