

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARZL-0146 |
| Location: | WALL TOWNSHIP POLICE | |
| Calibration File No.: | 00459 | Calib. Date: 03/07/2012 |
| Certification File No.: | 00460 | Cert. No.: 00010 |
| Linearity File No.: | 00461 | Cert. Date: 03/07/2012 |
| Solution File No.: | 00504 | Lin. Date: 03/07/2012 |
| Sequential File No.: | 00504 | Lin. No.: 00008 |
| | | Soln. Date: 07/17/2012 |
| | | Soln. No.: 00092 |
| | | File Date: 07/17/2012 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDXB S3-114 |
| Solution Control Lot: | 11H094 | Expires: 08/17/2013 |
| | | Bottle No.: 0255 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 11:36D | | |
| Control 1 EC | 0.100% | 11:37D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 11:37D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:38D | | |
| Control 2 EC | 0.098% | 11:38D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 11:38D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:39D | | |
| Control 3 EC | 0.099% | 11:40D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 11:40D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:40D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

[Handwritten Signature] #129

Badge No.: 129

Date: 07/17/2012