

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	01259	Calib. Date: 03/25/2015
Certification File No.:	01260	Calib. No.: 00025
Linearity File No.:	01261	Cert. Date: 03/25/2015
Solution File No.:	01273	Cert. No.: 00017
Sequential File No.:	01273	Lin. Date: 03/25/2015
		Lin. No.: 00017
		Soln. Date: 04/26/2015
		Soln. No.: 00210
		File Date: 04/26/2015

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	14D129	Expires: 04/29/2016
		Bottle No.: 0140

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:46D		
Control 1 EC	0.099%	05:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:47D		
Control 2 EC	0.099%	05:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:49D		
Control 3 EC	0.099%	05:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	05:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: _____



Badge No.: 178

Date: 04/26/2015