

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00886 Calib. Date: 03/25/2015 Calib. No.: 00018
Certification File No.: 00887 Cert. Date: 03/25/2015 Cert. No.: 00015
Linearity File No.: 00888 Lin. Date: 03/25/2015 Lin. No.: 00015
Solution File No.: 00895 Soln. Date: 04/26/2015 Soln. No.: 00175
Sequential File No.: 00895 File Date: 04/26/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 04/29/2016
Solution Control Lot: 14D129 Bottle No.: 0135

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:14D		
Control 1 EC	0.101%	05:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	05:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:15D		
Control 2 EC	0.101%	05:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	05:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:16D		
Control 3 EC	0.100%	05:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	05:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: _____



Badge No.: 178

Date: 04/26/2015