

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Wall Township Police Dept. 2700 Allaire Road Wall, NJ 07719	2. Contact: Ptln. Todd Kazanowsky	4. Date: 9/29/2011
	3. Phone Number: (732) 449-4500	5. County: Monmouth

6. Alcotest Instrument Serial Number: ARXA-0037	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A
---	---	---

9. Reason for Service:
Returned from Draeger

10. Comments:
Instrument was put back into service and a Calibration was performed on same with positive results.


11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

12. The above Instrument/Component is placed out of service pending further evaluation.

13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Marc Dennis #5925
Name & Badge Number (Print)


Signature

9/29/11
Date

RETURN AND REPAIR FORM



Customer Information

B: _____ S: _____

Company Name: Wall P.D., Wall NJ

Date Received: 09/13/11 Date given to service: _____

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810 7110 8610 7510 DT5000 7410 Upper-half 9510 7410 Whole

Serial Number: AR XA-0037
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>w/c w</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>w/c w</u>
<u>mp cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>w/c w</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>w/c w</u>

Repair Notes: Fuel cell ARCH-0731 replaced because of leakage, with Fuel cell ARCH-1027.
Cal w/ DC & ops check

Service Technician: [Signature] Date: 9-19-11

Packing Slip

Customer no Order No Order date
 150062190 10501286 09/20/2011

Packing slip no. Ship date
 80629334 09/20/2011

Please reference on inquiries

ship to
 150062190
 TOWNSHIP OF WALL
 ATTN: POLICE DEPARTMENT
 MONMOUTH COUNTY
 2700 ALLAIRE ROAD
 WALL, NJ 07719
 USA

bill to

Your Purchase Order Number

7110 WARRANTY: ARXA-0037
 Any questions? Please contact:
 MARA HERRERA

Customer
 TOWNSHIP OF WALL
 ATTN: POLICE DEPARTMENT
 MONMOUTH COUNTY
 2700 ALLAIRE ROAD
 WALL, NJ 07719
 USA

Your sales person

Don Pouliot
 Phone: 973-398-3228

Delivery terms
 FA
 FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Saver
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110: ARXA-0037
0010	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA
0020	1 EA	6808486 PLATES,7110--9510 Ordered / Back ordered : 0 / 0 EA

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 200337
 Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150062190 10501286 09/20/2011

Packing slip no. Ship date
80629334 09/20/2011

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 0 / 0 EA			
0040	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Wall Township Police Dept. 2700 Allaire Road Wall, NJ 07719		2. Contact: Ptln. Todd Kazanowsky		4. Date: 9/7/2011	
		3. Phone Number: (732) 449-4500		5. County: Monmouth	
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARXA-0037</p>		7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>		8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>	

9. Reason for Service:
At above location to perform a calibration and was advised of repeated control test failures, printed history of A.I.R's and data confirmed the same.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed several repeated control test failures due to low Electrochemical Sensor (EC) results. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

<p>Tpr. I Marc Dennis #5925 _____ Name & Badge Number (Print)</p>	<p><i>Tpr. I Marc Dennis</i> _____ Signature</p>	<p>9/7/11 _____ Date</p>
--	--	---