

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Wall Twp. Police Dept. 2700 Allaire Road Wall, NJ 07719		2. Contact: Ptlm. Todd Kazanowsky	4. Date: 4/11/2014
		3. Phone Number: (732) 449-4500	5. County: Monmouth
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARZL-0146</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>	

9. Reason for Service:
Returned from Draeger.

10. Comments:
Instrument was put back into service and a Calibration was performed on same with positive results.

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Sgt. Marc Dennis #5925

Name & Badge Number (Print)

SGT. M. Dennis 4/11/14

Signature Date

RETURN AND REPAIR FORM



Customer Information

B: _____ S: _____

Company Name: Wall Twp. P.D., Wall, NJ

Date Received: 03/18/14

Date given to service: 03/18/14

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR ZL-0146

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
MP Labor	Labor	10	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
8315075	Alcotest 7110 Motherboard	1	NC/W
6809458	IR Detector	1	NC/W

Repair Notes: Replaced defective Alcotest 7110 Motherboard, IR Detector, fuel cell and spacer plates.
Alcotest 7110 Motherboard was replaced due to the IRV not stabilizing and the IR Detector was replaced
due to unit giving "BLOWING NOT ALLOWED" error.

Fuel Cell = ARFB-0252

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 04/08/14

Packing Slip

Customer no Order No Order date
150062190 10692852 04/08/2014

Packing slip no. Ship date
80850038 04/08/2014
Please reference on inquiries

ship to
150062190
TOWNSHIP OF WALL
ATTN: POLICE DEPARTMENT
MONMOUTH COUNTY
2700 ALLAIRE ROAD
WALL, NJ 07719
USA

bill to

Your Purchase Order Number

PTLM. TODD KAZANOWSKY
Any questions? Please contact:
CHARLES ROBINSON II

Customer
TOWNSHIP OF WALL
ATTN: POLICE DEPARTMENT
MONMOUTH COUNTY
2700 ALLAIRE ROAD
WALL, NJ 07719
USA

Your sales person

Don Pouliot
Phone: 973-398-3228

Delivery terms
FA
FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110:ARZL-0146 RECEIVED 03/18/2014			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA			
0020	10 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0 / 0 H			

Draeger Safety Diagnostics Inc.
Accounting Address:
101 Technology Drive
Pittsburgh, PA 15275
Tel: (412) 788-5537
Fax: (412) 788-5598

Remit Wire Transfers To:
Citizens Bank
Acct. # 6209426615
Acct. Name: Draeger Safety Diagnostics Inc.
Transit Number: 036076150
Federal ID Number: 84-1600159

REMIT TO:
Draeger Safety Diagnostics Inc.
P.O. Box 200337
Pittsburgh, PA 15251-0337

Packing Slip

Customer no Order No Order date
150062190 10692852 04/08/2014

Packing slip no. Ship date
80850038 04/08/2014

Please reference on inquiries

Page 2 / 2


Item#	sh. Quant.	Part no. Description			
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			
0040	2 EA	6808486 PLATES, 7110--9510 Ordered / Back ordered : 0 / 0 EA			
0050	1 EA	8315075 PCB, ALCOTEST 7110 MKII US VERSION Ordered / Back ordered : 0 / 0 EA			
0060	1 EA	6809458 IR DETECTOR PCB Ordered / Back ordered : 0 / 0 EA			

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Wall Twp. Police Dept. 2700 Allaire Road Wall, NJ 07719	2. Contact: Ptlm. Todd Kazanowsky 3. Phone Number: (732) 449-4500	4. Date: 3/12/2014 5. County: Monmouth
6. Alcotest Instrument Serial Number: <p style="text-align: center;">ARZL-0146</p>	7. Simulator Component Serial Number: <p style="text-align: center;">N/A</p>	8. Temperature Probe Component Serial Number: <p style="text-align: center;">N/A</p>

9. Reason for Service:
Was advised by agency that instrument has been experiencing repeated control test failures, printed history of A.I.R's and data confirmed the same.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed several repeated control test failures due to low Electrochemical Sensor (EC) results. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.



Last Sequential File # 719

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

<p>Sgt. Marc Dennis #5925 Name & Badge Number (Print)</p>	<p><i>SGT. M. Dennis</i> Signature</p>	<p>3/12/14 Date</p>
--	--	--------------------------------