

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Wall Township Police Department 2700 Allaire Road Wall, NJ 07719		2. Contact: Todd Kazanowsky	4. Date: 9/15/2016
		3. Phone Number: 732-449-4500	5. County: Monmouth
6. Alcotest Instrument Serial Number: ARXA-0037	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A	

9. Reason for Service:
Alcotest instrument returned from outside evaluation or repair and placed back into service.

10. Comments:
See Draeger Return and Repair Form.

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Trooper II David Klimik #7040

Name & Badge Number (Print)

Trooper II D. Klimik #7040 9/15/16

Signature

Date

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Wall Twp. P.D., Wall, NJ

Date Received: 08/17/2016

Date given to service: 08/17/2016

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR XA-0037

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: November 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	1	NC/W
6808455	Fuel Cell	1	NC/W

Repair Notes: Replaced defecive fuel cell.

Fuel Cell = ARJC-1139

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 09/01/2016

Packing Slip

Customer no Order No Order date
 150062190 10853864 09/02/2016

Packing slip no. Ship date
 81031210 09/02/2016
 Please reference on inquiries

Ship to
 150062190
 TOWNSHIP OF WALL
 ATTN: POLICE DEPARTMENT
 MONMOUTH COUNTY
 2700 ALLAIRE ROAD
 WALL TOWNSHIP, NJ 07719-9570
 USA

Payer

Your Purchase Order Number

7110 ARXA-0037
 Any questions? Please contact:
 MARA BRACKENS

Customer
 TOWNSHIP OF WALL
 ATTN: POLICE DEPARTMENT
 MONMOUTH COUNTY
 2700 ALLAIRE ROAD
 WALL TOWNSHIP, NJ 07719-9570
 USA

Your sales person

CHRISTOPHER FAIRCHILD
 Phone: 866-385-5900
 Fax: 972-929-1260

Delivery terms
 PPA
 PREPAY AND ADD

Item#	sh. Quant.	Part no. Description			
		Waybill number: 1001921340360000771900677251643687 Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110 ARXA-0037			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA			

Draeger Safety Diagnostics Inc.
 Accounting Address:
 101 Technology Drive
 Pittsburgh, PA 15275
 Tel: (412) 788-5537
 Fax: (412) 788-5598

Remit Wire Transfers To:
 Citizens Bank
 Acct. # 6209426615
 Acct. Name: Draeger Safety Diagnostics Inc.
 Transit Number: 036076150
 Federal ID Number: 84-1600159

REMIT TO:
 Draeger Safety Diagnostics Inc.
 P.O. Box 536410
 Pittsburgh, PA 15253-5906

Packing Slip

Customer no Order No Order date
150062190 10853864 09/02/2016

Packing slip no. Ship date
81031210 09/02/2016

Please reference on inquiries

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0020	0.5 EA	4414166 HOURLY LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 EA			
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			