

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Wall Township Police Department 2700 Allaire Road Wall, NJ 07719</b>	2. Contact: <b>Todd Kazanowsky</b>	4. Date: <b>9/15/2016</b>
	3. Phone Number: <b>732-449-4500</b>	5. County: <b>Monmouth</b>
6. Alcotest Instrument Serial Number: <b>ARZL-0146</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>

9. Reason for Service:  
**Alcotest instrument returned from outside evaluation or repair and placed back into service.**

10. Comments:  
**See Draeger Return and Repair Form.**

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

<b>Trooper II David Klimik #7040</b> <small>Name &amp; Badge Number (Print)</small>	 <small>Signature</small>	<b>#7040 9/15/16</b> <small>Date</small>
--	--	---

# Packing Slip

Customer no    Order No    Order date  
150062190    10853906    09/06/2016

Packing slip no.    Ship date  
81031248    09/06/2016  
Please reference on inquiries

Ship to  
150062190  
TOWNSHIP OF WALL  
ATTN: POLICE DEPARTMENT/TODD KAZANOWSKY  
MONMOUTH COUNTY  
2700 ALLAIRE ROAD  
WALL TOWNSHIP, NJ 07719-9570  
USA

Payer

Your Purchase Order Number  
TODD KAZANOWSKY  
Any questions? Please contact:  
ROSEMARIE HUGHES

Customer  
TOWNSHIP OF WALL  
ATTN: POLICE DEPARTMENT  
MONMOUTH COUNTY  
2700 ALLAIRE ROAD  
WALL TOWNSHIP, NJ 07719-9570  
USA

Your sales person  
CHRISTOPHER FAIRCHILD  
Phone: 866-385-5900  
Fax: 972-929-1260

Delivery terms  
FA  
FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110:ARZL-0146 REC'D 8/18/16			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA			
0020	1 EA	4414166 HOURLY LABOR CHARGE Ordered / Back ordered : 0 / 0 EA			

**Draeger Safety Diagnostics Inc.**  
Accounting Address:  
101 Technology Drive  
Pittsburgh, PA 15275  
Tel: (412) 788-5537  
Fax: (412) 788-5598

**Remit Wire Transfers To:**  
Citizens Bank  
Acct. # 6209426615  
Acct. Name: Draeger Safety Diagnostics Inc.  
Transit Number: 036076150  
Federal ID Number: 84-1600159

**REMIT TO:**  
Draeger Safety Diagnostics Inc.  
P.O. Box 536410  
Pittsburgh, PA 15253-5906

# Packing Slip

Customer no    Order No    Order date  
150062190    10853906    09/06/2016

Packing slip no.    Ship date  
81031248    09/06/2016

Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			
0040	1 EA	6808374 GEARED ENGINE (MOTOR) Ordered / Back ordered : 0 / 0 EA			

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Wall Twp. P.D., Wall, NJ

Date Received: 08/19/2016

Date given to service: 08/19/2016

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR ZL-0146

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: November 2016

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

## Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	1	NC/W
6808455	Fuel Cell	1	NC/W
6808374	Geared Engine (Motor)	1	NC/W

Repair Notes: Replaced defective fuel cell and geared engine.

Fuel Cell = ARJC-2219

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 09/02/2016