

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00225	Calib. Date: 09/29/2010
Certification File No.:	00226	Cert. No.: 00007
Linearity File No.:	00227	Cert. Date: 09/29/2010
Solution File No.:	00268	Lin. No.: 00005
Sequential File No.:	00268	Lin. Date: 09/29/2010
		Soln. No.: 00045
		Soln. Date: 01/22/2011
		File Date: 01/22/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	10F079	Expires: 06/10/2012
		Bottle No.: 0397

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:39S		
Control 1 EC	0.102%	21:40S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:40S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:41S		
Control 2 EC	0.099%	21:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	21:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:42S		
Control 3 EC	0.100%	21:43S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COSTANTINI

First Name: JASON

MI: R

Signature: _____

Badge No.: 152

Date: 01/22/2011