

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00225	Calib. Date: 09/29/2010
Certification File No.:	00226	Cert. No.: 00007
Linearity File No.:	00227	Cert. Date: 09/29/2010
Solution File No.:	00271	Lin. No.: 00005
Sequential File No.:	00271	Lin. Date: 09/29/2010
		Soln. No.: 00046
		Soln. Date: 02/01/2011
		File Date: 02/01/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	10F079	Expires: 06/10/2012
		Bottle No.: 0382

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:27S		
Control 1 EC	0.101%	13:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:28S		
Control 2 EC	0.099%	13:29S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:30S		
Control 3 EC	0.099%	13:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 02/01/2011