

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00346	Calib. Date:	09/29/2010
Certification File No.:	00347	Cert. Date:	09/29/2010
Linearity File No.:	00348	Lin. Date:	09/29/2010
Solution File No.:	00383	Soln. Date:	02/15/2011
Sequential File No.:	00383	File Date:	02/15/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	10F079	Expires:	06/10/2012
		Bottle No.:	0379

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:37S		
Control 1 EC	-.---%	13:38S	21.4°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	13:38S	21.4°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*PTL #129*

Badge No.: 129

Date: 02/15/2011