

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00225 **Calib. Date:** 09/29/2010 **Calib. No.:** 00007
Certification File No.: 00226 **Cert. Date:** 09/29/2010 **Cert. No.:** 00005
Linearity File No.: 00227 **Lin. Date:** 09/29/2010 **Lin. No.:** 00005
Solution File No.: 00276 **Soln. Date:** 02/15/2011 **Soln. No.:** 00047
Sequential File No.: 00276 **File Date:** 02/15/2011

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 06/10/2012
Solution Control Lot: 10F079 **Bottle No.:** 0391

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:27S		
Control 1 EC	--%	13:27S	20.6°C	SIMULATOR TEMP. ERROR
Control 1 IR	--%	13:27S	20.6°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	--%	--:--		
Control 2 EC	--%	--:--	N/A	
Control 2 IR	--%	--:--	N/A	
Ambient Air Blank	--%	--:--		
Control 3 EC	--%	--:--	N/A	
Control 3 IR	--%	--:--	N/A	
Ambient Air Blank	--%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 02/15/2011