

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00346	Calib. Date: 09/29/2010
Certification File No.:	00347	Cert. No.: 00010
Linearity File No.:	00348	Cert. Date: 09/29/2010
Solution File No.:	00384	Lin. Date: 09/29/2010
Sequential File No.:	00384	Lin. No.: 00006
		Soln. Date: 02/16/2011
		Soln. No.: 00071
		File Date: 02/16/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	10F079	Expires: 06/10/2012
		Bottle No.: 0377

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:40S		
Control 1 EC	0.100%	10:41S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:42S		
Control 2 EC	0.098%	10:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:43S		
Control 3 EC	0.098%	10:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL K #129

Badge No.: 129

Date: 02/16/2011