

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00346	Calib. Date:	09/29/2010
Certification File No.:	00347	Cert. Date:	09/29/2010
Linearity File No.:	00348	Lin. Date:	09/29/2010
Solution File No.:	00387	Soln. Date:	03/03/2011
Sequential File No.:	00387	File Date:	03/03/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	10L084	Expires:	12/16/2012
		Bottle No.:	1221

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:26S		
Control 1 EC	0.103%	13:27S	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.108%	13:27S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	13:28S		
Control 2 EC	0.100%	13:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:29S		
Control 3 EC	0.100%	13:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:31S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

PTL K #129

Badge No.: 129

Date: 03/03/2011