

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE	Calib. Date:	09/29/2010
Calibration File No.:	00225	Cert. No.:	00007
Certification File No.:	00226	Cert. Date:	09/29/2010
Linearity File No.:	00227	Lin. No.:	00005
Solution File No.:	00283	Soln. Date:	03/03/2011
Sequential File No.:	00283	File Date:	03/03/2011

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDAB-0014
Control Solution %:	0.100%	Expires:	06/10/2012	Bottle No.:	0396
Solution Control Lot:	10F079				

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:15S		
Control 1 EC	0.101%	13:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:17S		
Control 2 EC	0.098%	13:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:18S		
Control 3 EC	0.099%	13:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*Plc. Kazanowski*

Badge No.: 129

Date: 03/03/2011