

# Calibrating Unit New Standard Solution Report

|                         |                       |                         |
|-------------------------|-----------------------|-------------------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.: ARXA-0037   |
| Location:               | WALL TOWNSHIP POLICE  |                         |
| Calibration File No.:   | 00346                 | Calib. Date: 09/29/2010 |
| Certification File No.: | 00347                 | Calib. No.: 00010       |
| Linearity File No.:     | 00348                 | Cert. Date: 09/29/2010  |
| Solution File No.:      | 00391                 | Cert. No.: 00006        |
| Sequential File No.:    | 00391                 | Lin. Date: 09/29/2010   |
|                         |                       | Lin. No.: 00006         |
|                         |                       | Soln. Date: 03/22/2011  |
|                         |                       | Soln. No.: 00075        |
|                         |                       | File Date: 03/22/2011   |

|                       |        |                  |                       |
|-----------------------|--------|------------------|-----------------------|
| Calibrating Unit:     | WET    | Model No.: CU-34 | Serial No.: DDAB-0015 |
| Control Solution %:   | 0.100% |                  | Expires: 12/16/2012   |
| Solution Control Lot: | 10L084 |                  | Bottle No.: 1211      |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 10:17S |                |                     |
| Control 1 EC      | 0.102% | 10:18S | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.101% | 10:18S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:19S |                |                     |
| Control 2 EC      | 0.098% | 10:19S | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.100% | 10:19S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:20S |                |                     |
| Control 3 EC      | 0.099% | 10:21S | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 10:21S | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:22S |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

Badge No.: 129

Date: 03/22/2011