

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARZL-0146 |
| Location: | WALL TOWNSHIP POLICE | |
| Calibration File No.: | 00225 | Calib. Date: 09/29/2010 |
| Certification File No.: | 00226 | Cert. No.: 00007 |
| Linearity File No.: | 00227 | Cert. Date: 09/29/2010 |
| Solution File No.: | 00290 | Lin. No.: 00005 |
| Sequential File No.: | 00290 | Lin. Date: 09/29/2010 |
| | | Soln. No.: 00051 |
| | | Soln. Date: 03/22/2011 |
| | | File Date: 03/22/2011 |

| | | | |
|-----------------------|--------|------------------|-----------------------|
| Calibrating Unit: | 1 | Model No.: CU-34 | Serial No.: DDAB-0014 |
| Control Solution %: | 0.100% | | Expires: 12/16/2012 |
| Solution Control Lot: | 10L084 | | Bottle No.: 1216 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 10:07S | | |
| Control 1 EC | 0.103% | 10:07S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.102% | 10:07S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:08S | | |
| Control 2 EC | 0.099% | 10:09S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 10:09S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:10S | | |
| Control 3 EC | 0.099% | 10:10S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 10:10S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 10:11S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 03/22/2011