

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00393	Calib. Date:	03/29/2011
Certification File No.:	00394	Cert. No.:	00011
Linearity File No.:	00395	Cert. Date:	03/29/2011
Solution File No.:	00396	Lin. Date:	03/29/2011
Sequential File No.:	00396	Soln. Date:	03/29/2011
		File Date:	03/29/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	10L084	Expires:	12/16/2012
		Bottle No.:	1207

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:51S	12:51 <i>JJS</i>	
Control 1 EC	0.101%	11:52S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:53S		
Control 2 EC	0.099%	11:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:54S		
Control 3 EC	0.100%	11:55S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:56S	12:56 <i>JJS</i>	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDWS P2-283 *(JJS)*

Changed By:

Last Name: SNYDER

First Name: THOMAS

MI: J.

Signature: *THOMAS J. Snyder #5792*

Badge No.: 5792

Date: 03/29/2011