

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00295 **Calib. Date:** 03/29/2011 **Calib. No.:** 00008
Certification File No.: 00296 **Cert. Date:** 03/29/2011 **Cert. No.:** 00006
Linearity File No.: 00297 **Lin. Date:** 03/29/2011 **Lin. No.:** 00006
Solution File No.: 00298 **Soln. Date:** 03/29/2011 **Soln. No.:** 00053
Sequential File No.: 00298 **File Date:** 03/29/2011

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 12/16/2012
Solution Control Lot: 10L084 **Bottle No.:** 1201

Function	Result	Time	Temperature	Comment(s) or Error(s)
	%BAC	HH:MM	Simulator (°C)	
Ambient Air Blank	0.000%	12:25S	13:25b <i>JYS</i>	
Control 1 EC	0.100%	12:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26S		
Control 2 EC	0.098%	12:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:28S		
Control 3 EC	0.099%	12:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:29S	13:27b <i>JYS</i>	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDWS P2-284 *JYS*

Changed By:

Last Name: SNYDER

First Name: THOMAS

MI: J.

Signature: *THOMAS J. Snyder #5792*

Badge No.: 5792

Date: 03/29/2011