

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00393	Calib. Date: 03/29/2011
Certification File No.:	00394	Calib. No.: 00011
Linearity File No.:	00395	Cert. Date: 03/29/2011
Solution File No.:	00399	Cert. No.: 00007
Sequential File No.:	00399	Lin. Date: 03/29/2011
		Lin. No.: 00007
		Soln. Date: 04/01/2011
		Soln. No.: 00077
		File Date: 04/01/2011

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXB S3-114
Control Solution %:	0.100%		Expires: 12/16/2012
Solution Control Lot:	10L084		Bottle No.: 1222

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:30S		
Control 1 EC	0.099%	08:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:31S		
Control 2 EC	0.097%	08:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32S		
Control 3 EC	0.097%	08:33S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 04/01/2011