

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00393	Calib. Date:	03/29/2011
Certification File No.:	00394	Cert. No.:	00011
Linearity File No.:	00395	Cert. Date:	03/29/2011
Solution File No.:	00413	Lin. Date:	03/29/2011
Sequential File No.:	00413	Lin. No.:	00007
		Soln. Date:	05/02/2011
		Soln. No.:	00079
		File Date:	05/02/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	10L084	Expires:	12/16/2012
		Bottle No.:	0756

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:12D		
Control 1 EC	0.099%	16:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:14D		
Control 2 EC	0.097%	16:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:15D		
Control 3 EC	0.098%	16:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: COSTANTINI

First Name: JASON

MI: R

Signature: _____

Badge No.: 152

Date: 05/02/2011