

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00295	Calib. Date:	03/29/2011
Certification File No.:	00296	Cert. No.:	00008
Linearity File No.:	00297	Cert. Date:	03/29/2011
Solution File No.:	00316	Lin. Date:	03/29/2011
Sequential File No.:	00316	Lin. No.:	00006
		Soln. Date:	05/02/2011
		Soln. No.:	00055
		File Date:	05/02/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	10L084	Expires:	12/16/2012
		Bottle No.:	0761

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:17D		
Control 1 EC	0.099%	15:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	15:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:18D		
Control 2 EC	0.097%	15:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19D		
Control 3 EC	0.097%	15:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: COSTANTINI

First Name: JASON

MI: R

Signature: \_\_\_\_\_

Badge No.: 152

Date: 05/02/2011