

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00295	Calib. Date:	03/29/2011
Certification File No.:	00296	Cert. No.:	00008
Linearity File No.:	00297	Lin. Date:	03/29/2011
Solution File No.:	00322	Lin. No.:	00006
Sequential File No.:	00322	Soln. Date:	05/16/2011
		Soln. No.:	00056
		File Date:	05/16/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	10L084	Expires:	12/16/2012
		Bottle No.:	1204

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:51D		
Control 1 EC	---%	11:51D	20.9°C	SIMULATOR TEMP. ERROR
Control 1 IR	---%	11:51D	20.9°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	---%	--:--		
Control 2 EC	---%	--:--	N/A	
Control 2 IR	---%	--:--	N/A	
Ambient Air Blank	---%	--:--		
Control 3 EC	---%	--:--	N/A	
Control 3 IR	---%	--:--	N/A	
Ambient Air Blank	---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PK 129

Badge No.: 129

Date: 05/16/2011