

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00393	Calib. Date: 03/29/2011
Certification File No.:	00394	Calib. No.: 00011
Linearity File No.:	00395	Cert. Date: 03/29/2011
Solution File No.:	00424	Cert. No.: 00007
Sequential File No.:	00424	Lin. Date: 03/29/2011
		Lin. No.: 00007
		Soln. Date: 05/18/2011
		Soln. No.: 00080
		File Date: 05/18/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 1214

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:27D		
Control 1 EC	0.099%	07:27D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:28D		
Control 2 EC	0.098%	07:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:29D		
Control 3 EC	0.097%	07:30D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: \_\_\_\_\_

 #162

Badge No.: 162

Date: 05/18/2011