Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE

Calibration File No.:	00295	Calib. Date:	03/29/2011	Calib. No.:	00008
Certification File No.:	00296	Cert. Date:	03/29/2011	Cert. No.:	00006
Linearity File No.:	00297	Lin. Date:	03/29/2011	Lin. No.:	00006
Solution File No.:	00326	Soln. Date:	05/18/2011	Soln. No.:	00059
Sequential File No.:	00326	File Date:	05/18/2011		

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 12/16/2012
Solution Control Lot: 10L084 Bottle No.: 1224

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:04D		
Control 1 EC	0.099%	16:05D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	16:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:05D		
Control 2 EC	0.095%	16:06D	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.093%	16:06D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	16:07D		
Control 3 EC	0.096%	16:07D	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.092%	16:07D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	16:08D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: COSTANTINI First Name: JASON MI: R
Signature: 152 Badge No.: 152
Date: 05/18/2011