

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
 Calibration File No.: 00393 Calib. Date: 03/29/2011 Calib. No.: 00011
 Certification File No.: 00394 Cert. Date: 03/29/2011 Cert. No.: 00007
 Linearity File No.: 00395 Lin. Date: 03/29/2011 Lin. No.: 00007
 Solution File No.: 00426 Soln. Date: 05/21/2011 Soln. No.: 00081
 Sequential File No.: 00426 File Date: 05/21/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 12/16/2012
Solution Control Lot: 10L084 Bottle No.: 1217

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:53D		
Control 1 EC	0.100%	00:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:55D		
Control 2 EC	0.099%	00:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:56D		
Control 3 EC	0.098%	00:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

P. J. Cadigan #162

Badge No.: 162

Date: 05/21/2011