

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00393	Calib. Date:	03/29/2011
Certification File No.:	00394	Cert. No.:	00007
Linearity File No.:	00395	Lin. Date:	03/29/2011
Solution File No.:	00437	Soln. No.:	00083
Sequential File No.:	00437	File Date:	06/14/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	10L084	Expires:	12/16/2012
		Bottle No.:	1210

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:17D		
Control 1 EC	0.100%	04:17D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:18D		
Control 2 EC	0.098%	04:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:19D		
Control 3 EC	0.098%	04:20D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

 #162

Badge No.: 162

Date: 06/14/2011