

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00393	Calib. Date:	03/29/2011
Certification File No.:	00394	Cert. Date:	03/29/2011
Linearity File No.:	00395	Lin. Date:	03/29/2011
Solution File No.:	00442	Soln. Date:	06/20/2011
Sequential File No.:	00442	File Date:	06/20/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	10L084	Expires:	12/16/2012
		Bottle No.:	1219

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:42D		
Control 1 EC	0.099%	02:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:43D		
Control 2 EC	0.097%	02:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:45D		
Control 3 EC	0.097%	02:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 06/20/2011