

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00393	Calib. Date: 03/29/2011
Certification File No.:	00394	Cert. No.: 00011
Linearity File No.:	00395	Cert. Date: 03/29/2011
Solution File No.:	00448	Lin. Date: 03/29/2011
Sequential File No.:	00448	Soln. Date: 06/29/2011
		Soln. No.: 00086
		File Date: 06/29/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 1213

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:00D		
Control 1 EC	0.099%	07:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:02D		
Control 2 EC	0.098%	07:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:03D		
Control 3 EC	0.098%	07:04D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

Badge No.: 162

Date: 06/29/2011