

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00295	Calib. No.: 00008
Certification File No.:	00296	Cert. No.: 00006
Linearity File No.:	00297	Lin. No.: 00006
Solution File No.:	00352	Soln. No.: 00063
Sequential File No.:	00352	File Date: 07/01/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 1203

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:54D		
Control 1 EC	0.099%	10:54D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	10:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:55D		
Control 2 EC	0.097%	10:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	10:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56D		
Control 3 EC	0.096%	10:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	10:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WOLTER

First Name: STEVEN

MI: R

Signature: \_\_\_\_\_



Badge No.: 114

Date: 07/01/2011