

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00393	Calib. Date: 03/29/2011
Certification File No.:	00394	Cert. No.: 00011
Linearity File No.:	00395	Cert. Date: 03/29/2011
Solution File No.:	00458	Lin. No.: 00007
Sequential File No.:	00458	Lin. Date: 03/29/2011
		Soln. No.: 00090
		Soln. Date: 07/07/2011
		File Date: 07/07/2011

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXB S3-114
Control Solution %:	0.100%		Expires: 12/16/2012
Solution Control Lot:	10L084		Bottle No.: 0722

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:08D		
Control 1 EC	0.099%	11:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09D		
Control 2 EC	0.097%	11:10D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:10D		
Control 3 EC	0.097%	11:11D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	11:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLTER

First Name: STEVEN

MI: R

Signature: _____

SGT. S.A. Wolter 114

Badge No.: 114

Date: 07/07/2011