

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00393	Calib. Date:	03/29/2011
Certification File No.:	00394	Cert. No.:	00007
Linearity File No.:	00395	Lin. Date:	03/29/2011
Solution File No.:	00465	Soln. No.:	00091
Sequential File No.:	00465	File Date:	07/21/2011

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXB S3-114
Control Solution %:	0.100%			Expires:	12/16/2012
Solution Control Lot:	10L084			Bottle No.:	0762

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:12D		
Control 1 EC	0.099%	07:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	07:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:13D		
Control 2 EC	0.096%	07:14D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:14D		
Control 3 EC	0.097%	07:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: \_\_\_\_\_

Badge No.: 162

Date: 07/21/2011