

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00295 Calib. Date: 03/29/2011 Calib. No.: 00008
Certification File No.: 00296 Cert. Date: 03/29/2011 Cert. No.: 00006
Linearity File No.: 00297 Lin. Date: 03/29/2011 Lin. No.: 00006
Solution File No.: 00365 Soln. Date: 07/21/2011 Soln. No.: 00065
Sequential File No.: 00365 File Date: 07/21/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 12/16/2012
Solution Control Lot: 10L084 Bottle No.: 1220

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 07:06D | | |
| Control 1 EC | 0.099% | 07:07D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 07:07D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:08D | | |
| Control 2 EC | 0.096% | 07:08D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 07:08D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:09D | | |
| Control 3 EC | 0.096% | 07:10D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.098% | 07:10D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:11D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CADIGAN

First Name: JAMES

MI: R

Signature: _____

James Cadigan
#162

Badge No.: 162

Date: 07/21/2011