

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00393	Calib. No.: 00011
Certification File No.:	00394	Cert. No.: 00007
Linearity File No.:	00395	Lin. No.: 00007
Solution File No.:	00472	Soln. No.: 00093
Sequential File No.:	00472	File Date: 07/31/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-114
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 0766

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:53D		
Control 1 EC	0.098%	03:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:54D		
Control 2 EC	0.096%	03:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:55D		
Control 3 EC	0.097%	03:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	03:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:56D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLTER

First Name: STEVEN

MI: R

Signature: _____

SGT S. Wolter #114

Badge No.: 114

Date: 07/31/2011