

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00295	Calib. Date: 03/29/2011
Certification File No.:	00296	Cert. No.: 00008
Linearity File No.:	00297	Cert. Date: 03/29/2011
Solution File No.:	00385	Lin. No.: 00006
Sequential File No.:	00385	Soln. Date: 08/13/2011
		Soln. No.: 00067
		File Date: 08/13/2011

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXB S3-113
Control Solution %:	0.100%		Expires: 12/16/2012
Solution Control Lot:	10L084		Bottle No.: 0765

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:21D		
Control 1 EC	0.098%	13:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:23D		
Control 2 EC	0.097%	13:23D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:24D		
Control 3 EC	0.097%	13:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated *NEW SOLUTION* in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LISTER

First Name: EDWARD

MI: J

Signature: _____



Badge No.: 142

Date: 08/13/2011