

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C		Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE		
Calibration File No.: 00295	Calib. Date: 03/29/2011	Calib. No.: 00008
Certification File No.: 00296	Cert. Date: 03/29/2011	Cert. No.: 00006
Linearity File No.: 00297	Lin. Date: 03/29/2011	Lin. No.: 00006
Solution File No.: 00387	Soln. Date: 08/17/2011	Soln. No.: 00068
Sequential File No.: 00387	File Date: 08/17/2011	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDXB S3-113
Control Solution %: 0.100%		Expires: 12/16/2012
Solution Control Lot: 10L084		Bottle No.: 0753

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:17D		
Control 1 EC	0.099%	08:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:19D		
Control 2 EC	0.097%	08:19D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:20D		
Control 3 EC	0.098%	08:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:21D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

[Handwritten Signature] #129

Badge No.: 129

Date: 08/17/2011