

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00295	Calib. Date: 03/29/2011
Certification File No.:	00296	Calib. No.: 00008
Linearity File No.:	00297	Cert. Date: 03/29/2011
Solution File No.:	00394	Cert. No.: 00006
Sequential File No.:	00394	Lin. Date: 03/29/2011
		Lin. No.: 00006
		Soln. Date: 08/23/2011
		Soln. No.: 00069
		File Date: 08/23/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 0752

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:34D		
Control 1 EC	0.099%	22:34D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:35D		
Control 2 EC	0.097%	22:36D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	22:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:36D		
Control 3 EC	0.097%	22:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WOLTER

First Name: STEVEN

MI: -

Signature: _____

Steven J. Wolter #114

Badge No.: 114

Date: 08/23/2011