

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00295	Calib. Date: 03/29/2011
Certification File No.:	00296	Calib. No.: 00008
Linearity File No.:	00297	Cert. Date: 03/29/2011
Solution File No.:	00397	Cert. No.: 00006
Sequential File No.:	00397	Lin. Date: 03/29/2011
		Lin. No.: 00006
		Soln. Date: 09/02/2011
		Soln. No.: 00070
		File Date: 09/02/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXB S3-113
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 0775

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:45D		
Control 1 EC	0.098%	09:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47D		
Control 2 EC	0.096%	09:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48D		
Control 3 EC	0.097%	09:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

 #129

Badge No.: 129

Date: 09/02/2011