

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARZL-0146  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 00402 Calib. Date: 09/07/2011 Calib. No.: 00009  
Certification File No.: 00403 Cert. Date: 09/07/2011 Cert. No.: 00007  
Linearity File No.: 00404 Lin. Date: 09/07/2011 Lin. No.: 00007  
Solution File No.: 00414 Soln. Date: 09/21/2011 Soln. No.: 00072  
Sequential File No.: 00414 File Date: 09/21/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113  
Control Solution %: 0.100% Expires: 03/31/2013  
Solution Control Lot: 11C088 Bottle No.: 0510

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:13D		
Control 1 EC	0.099%	11:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:15D		
Control 2 EC	0.098%	11:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:16D		
Control 3 EC	0.098%	11:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:17D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NASH

First Name: STEVEN

MI: P

Signature:  #157

Badge No.: 157

Date: 09/21/2011