

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00402	Calib. Date: 09/07/2011
Certification File No.:	00403	Calib. No.: 00009
Linearity File No.:	00404	Cert. Date: 09/07/2011
Solution File No.:	00416	Cert. No.: 00007
Sequential File No.:	00416	Lin. Date: 09/07/2011
		Lin. No.: 00007
		Soln. Date: 09/29/2011
		Soln. No.: 00073
		File Date: 09/29/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 0760

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:02D		
Control 1 EC	0.100%	12:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:03D		
Control 2 EC	0.099%	12:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:05D		
Control 3 EC	0.098%	12:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL # DOXY P2-393 (20)

Changed By:

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: T.M. I [Signature] #5925

Badge No.: 5925

Date: 09/29/2011