

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00402	Calib. Date: 09/07/2011
Certification File No.:	00403	Calib. No.: 00009
Linearity File No.:	00404	Cert. Date: 09/07/2011
Solution File No.:	00421	Cert. No.: 00007
Sequential File No.:	00421	Lin. Date: 09/07/2011
		Lin. No.: 00007
		Soln. Date: 10/14/2011
		Soln. No.: 00074
		File Date: 10/14/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	10L084	Expires: 12/16/2012
		Bottle No.: 0754

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:09D		
Control 1 EC	0.101%	09:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:10D		
Control 2 EC	0.098%	09:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:12D		
Control 3 EC	0.098%	09:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:13D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: 

Badge No.: 129

Date: 10/14/2011