

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00402	Calib. Date:	09/07/2011
Certification File No.:	00403	Cert. Date:	09/07/2011
Linearity File No.:	00404	Lin. Date:	09/07/2011
Solution File No.:	00428	Soln. Date:	11/04/2011
Sequential File No.:	00428	File Date:	11/04/2011
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	10L084	Expires:	12/16/2012
		Bottle No.:	0774

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:16S		
Control 1 EC	0.100%	07:17S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:17S		
Control 2 EC	0.098%	07:18S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:19S		
Control 3 EC	0.098%	07:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL #129

Badge No.: 129

Date: 11/04/2011