

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00402	Calib. Date: 09/07/2011 Calib. No.: 00009
Certification File No.:	00403	Cert. Date: 09/07/2011 Cert. No.: 00007
Linearity File No.:	00404	Lin. Date: 09/07/2011 Lin. No.: 00007
Solution File No.:	00435	Soln. Date: 11/18/2011 Soln. No.: 00076
Sequential File No.:	00435	File Date: 11/18/2011
Calibrating Unit:	WET	Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %:	0.100%	Expires: 12/16/2012
Solution Control Lot:	10L084	Bottle No.: 0769

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:35S		
Control 1 EC	-.---%	08:36S	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	08:36S	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____



Badge No.: 129

Date: 11/18/2011