

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00494	Calib. Date: 09/29/2011
Certification File No.:	00495	Calib. No.: 00014
Linearity File No.:	00496	Cert. Date: 09/29/2011
Solution File No.:	00510	Cert. No.: 00008
Sequential File No.:	00510	Lin. Date: 09/29/2011
		Lin. No.: 00008
		Soln. No.: 00104
		File Date: 12/02/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	11H094	Expires: 08/17/2013
		Bottle No.: 0271

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:08S		
Control 1 EC	0.098%	13:08S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:09S		
Control 2 EC	0.098%	13:10S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:10S		
Control 3 EC	0.099%	13:11S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL. K #129

Badge No.: 129

Date: 12/02/2011