

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARZL-0146
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00402	Calib. Date: 09/07/2011
Certification File No.:	00403	Calib. No.: 00009
Linearity File No.:	00404	Cert. Date: 09/07/2011
Solution File No.:	00439	Cert. No.: 00007
Sequential File No.:	00439	Lin. Date: 09/07/2011
		Lin. No.: 00007
		Soln. Date: 12/02/2011
		Soln. No.: 00078
		File Date: 12/02/2011
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0015
Solution Control Lot:	11H094	Expires: 08/17/2013
		Bottle No.: 0266

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:10S		
Control 1 EC	0.099%	13:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:11S		
Control 2 EC	0.097%	13:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:13S		
Control 3 EC	0.098%	13:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_



Badge No.: 129

Date: 12/02/2011