

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00402	Calib. Date:	09/07/2011
Certification File No.:	00403	Cert. Date:	09/07/2011
Linearity File No.:	00404	Lin. Date:	09/07/2011
Solution File No.:	00453	Soln. Date:	02/01/2012
Sequential File No.:	00453	File Date:	02/01/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	11C088	Expires:	03/31/2013
		Bottle No.:	0523

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:49S		
Control 1 EC	0.100%	11:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:51S		
Control 2 EC	0.098%	11:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52S		
Control 3 EC	0.098%	11:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL #129

Badge No.: 129

Date: 02/01/2012