

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0037
Location:	WALL TOWNSHIP POLICE	
Calibration File No.:	00494	Calib. Date: 09/29/2011
Certification File No.:	00495	Cert. No.: 00014
Linearity File No.:	00496	Cert. Date: 09/29/2011
Solution File No.:	00550	Lin. No.: 00008
Sequential File No.:	00550	Lin. Date: 09/29/2011
		Soln. No.: 00108
		Soln. Date: 02/15/2012
		File Date: 02/15/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAB-0014
Solution Control Lot:	11C088	Expires: 03/31/2013
		Bottle No.: 0515

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:01S		
Control 1 EC	0.100%	11:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:02S		
Control 2 EC	0.101%	11:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:03S		
Control 3 EC	0.099%	11:04S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PLC. KA 129

Badge No.: 129

Date: 02/15/2012