

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	00402	Calib. Date:	09/07/2011
Certification File No.:	00403	Cert. No.:	00009
Linearity File No.:	00404	Cert. Date:	09/07/2011
Solution File No.:	00454	Lin. No.:	00007
Sequential File No.:	00454	Soln. Date:	02/15/2012
		Soln. No.:	00082
		File Date:	02/15/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	11C088	Expires:	03/31/2013
		Bottle No.:	0518

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:09S		
Control 1 EC	0.103%	11:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:11S		
Control 2 EC	0.099%	11:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:12S		
Control 3 EC	0.100%	11:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:14S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: 

Badge No.: 129

Date: 02/15/2012