

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00494 Calib. Date: 09/29/2011 Calib. No.: 00014
Certification File No.: 00495 Cert. Date: 09/29/2011 Cert. No.: 00008
Linearity File No.: 00496 Lin. Date: 09/29/2011 Lin. No.: 00008
Solution File No.: 00553 Soln. Date: 02/28/2012 Soln. No.: 00109
Sequential File No.: 00553 File Date: 02/28/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014
Control Solution %: 0.100% Expires: 03/31/2013
Solution Control Lot: 11C088 Bottle No.: 0520

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 11:58S | | |
| Control 1 EC | 0.100% | 11:59S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 11:59S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:59S | | |
| Control 2 EC | 0.099% | 12:00S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 12:00S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:00S | | |
| Control 3 EC | 0.099% | 12:01S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 12:01S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:02S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL #129

Badge No.: 129

Date: 02/28/2012