

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 00402 Calib. Date: 09/07/2011 Calib. No.: 00009
Certification File No.: 00403 Cert. Date: 09/07/2011 Cert. No.: 00007
Linearity File No.: 00404 Lin. Date: 09/07/2011 Lin. No.: 00007
Solution File No.: 00457 Soln. Date: 02/28/2012 Soln. No.: 00083
Sequential File No.: 00457 File Date: 02/28/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 03/31/2013
Solution Control Lot: 11C088 Bottle No.: 0524

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:59S		
Control 1 EC	0.100%	12:00S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:01S		
Control 2 EC	0.098%	12:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02S		
Control 3 EC	0.099%	12:03S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL #129

Badge No.: 129

Date: 02/28/2012